

# Intoxilyzer 8000 Fields

Case ID #

Subject Name: Last, First, MI

Subject Date of Birth

Subject Driver's License #

State of Issuance

Arresting Officer

Instrument Operator

Operator Number

Deprivation Period Began

## Dry Gas Cylinder Information

- a. The cylinders are purchased and tested by KDHE prior to being furnished to any law enforcement agency. Every cylinder is verified and must read within the acceptable external standard check range, 0.075 to 0.085 g/210 L, inclusive, or the cylinder is rejected.
- b. The cylinders have a shelf life of approximately two years. The cylinders are replaced prior to their expiration date.
- c. The cylinders are furnished at site inspections or mailed at the agency's request.
- d. They are made of steel and are non-refillable. They can be recycled or land filled after the pressure is released. Your Agency or Device Custodian has an extractor device used to ventilate tanks to remove the pressure which allows for safe disposal.
- e. The cylinders contain ethanol (ethyl alcohol) and nitrogen. The nitrogen is used as a propellant to help carry the ethanol into the instrument.
- f. The cylinders are considered a hazardous material due to the pressure in the cylinder not the contents. In order to be able to ship the cylinders you must be Department of Transportation (DOT) Hazardous Materials Shipping Certified. KDHE Breath Alcohol Program staff has obtained the required training to be able to ship these cylinders.
- g. When shipping an instrument for repair the dry gas cylinder **NEEDS TO BE REMOVED** to prevent fines and/or added expenses for shipping.

# Completed Test Printout

KDHE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002124  
06/12/2007 11:20

Case Iden No = 1111  
Sub Name = DOE, JOHN  
Sub DOB = 10/22/1950  
Sub Driv Lic = KS123456  
State of Issue = KS  
Arrest Officer = JANE DOE  
Inst Operator = JANE DOE  
Operator Cert No. = 0001  
Deprivation Period Began = 11:00

Test	g/210L BrAC	Time
Air Blank	0.000	11:22
Ext Std Check	0.078	11:22
Air Blank	0.000	11:23
Subject Test	0.079	11:23
Air Blank	0.000	11:24

---

JANE DOE

## Possible Cause:

This is a properly completed subject test print out.

## Potential Solution:

None.